

measure. analyze. innovate.

3-Component Force Link

25 x 25 x 30 mm up to 2 kN

Miniature force link for measuring the three orthogonal components of a dynamic or quasistatic force acting in an arbitrary direction.

- Very extended measuring range
- High rigidity
- Minimal cross talk
- Calibrated force link
- Simple installation
- Multipol connector

Description

The force sensor is mounted under preload between two plates and, therefore can measure compression and tensile forces. The piezo element yields an electric charge which is proportional to the force. This is picked off by electrodes and transferred via a connector. The charge signal is fed via a screened cable to a charge amplifier, which converts it into a proportional output voltage. An appropriate evaluation circuit can record and further process the measurand.

Thanks to their high rigidity they cover a wide frequency range. The elastic behaviour of the test object is practically not influenced. Quasistatic measurements, are possible, too.

The sensor is mounted ground-isolated. This largely eliminates ground loop problems.


The force link is supplied calibrated. After correct installation, it is immediately ready for use without re-calibration.

Application

The force links can measure the 3 orthogonal force components easily, directly and precisely, because these sensors feature an inherently low cross talk.

Application examples

- Cutting forces
- Impact forces
- Determination of coefficients of friction



Technical Data

		NI	4 000 4 000
Range	F _x , F _y	N	-1 000 1 000
(Without moments if e.g. four			
force links are mounted into			
one force plate)			
Range	F _x , F _y	N	-200 200
(Example with point of force			
application F _{x,y} 12 mm above			
top plate)			
Range	Fz	Ν	-2 000 2 000
(Point of force application F_z			
centric)			
Overload		%	10
Calibrated range	F _x , F _y	Ν	0 600; 0 60
(Point of force application $F_{x,y}$			
7,5 mm below top plate surface)			
Calibrated range	Fz	Ν	0 2000; 0 200
(Point of force application F_z			
centric)			
Max. moments			
$(M_z = 0; F_z = 0)$	M _{x,y}	N∙m	-5/5
$(M_{x,y} = 0; F_z = 0)$	Mz	N∙m	-4/4
Threshold		Ν	<0,01

Page 1/3

This information corresponds to the current state of knowledge. Kistler reserves the right to make technical changes. Liability for consequential damage resulting from the use of Kistler products is excluded.

©2004, Kistler Instrumente AG, PO Box, Eulachstr. 22, CH-8408 Winterthur Tel +41 52 224 11 11, Fax 224 14 14, info@kistler.com, www.kistler.com

Туре 9317В

measure. analyze. innovate.

Sensitivity	F _x , F _y	pC/N	≈–26
	Fz	pC/N	≈–11
Linearity, each axis		% FSO	≤±0,5
Hysteresis, each axis		% FSO	≤0,5
Cross talk	$F_z \rightarrow F_x, F_y$	%	≤±1
(Cross talk $F_x, F_y \rightarrow F_z$ is	$F_x \leftrightarrow F_y$	%	≤±3
≤±3% if e.g. four force links	$F_x, F_y \rightarrow F_z$	%	≤±4
are mounted in to one force			
plate)			
Natural frequency	f _n (x)	kHz	≈5
	f _n (y)	kHz	≈5
	f _n (z)	kHz	≈21
Operating temperature range		°C	-50 80
Insulation resistance		Ω	≥10 ¹³
Ground insulation		Ω	>108
Capacitance, each channel		pF	≈39
Connector			3 pole
			M8 x 0,75
Weight		g	85

Multicomponent Force Plate with 4 Force Links

Four force links can be mounted together with a common top plate. Such we gain a multicomponent force plate.





Fig. 1: Force plate with 4 force links Type 9317B



This information corresponds to the current state of knowledge. Kistler reserves the right to make technical changes. Liability for consequential damage resulting from the use of Kistler products is excluded.

©2004, Kistler Instrumente AG, PO Box, Eulachstr. 22, CH-8408 Winterthur Tel +41 52 224 11 11, Fax 224 14 14, info@kistler.com, www.kistler.com

Page 2/3

measure. analyze. innovate.

Mounting

The contact surfaces that transfer the forces onto the force link must be flat, rigid and clean.

When four force links are used to construct a force plate, they must be machined to the same level.

The force links can be fastened either from outside with four screws M4 in each case or from the center again in each case with four screws M3.

The screws must be tightened sufficiently so that even with maximum force exerted, no gap occurs between the contact surfaces.




Fig. 3: Mounting example: Force limited vibration testing

Parallel Switching

Several force links of identical sensitivities can be paralleled directly. The charge amplifier connected then gives an output voltage which corresponds to the sum of all forces acting.

This is a great advantage when building force plates with which only the 3 components of the resulting force must be measured.

Electronics

Besides the force sensors, a 3-component force measuring system also requires 3 charge amplifiers, which convert the electrical charge signals of the sensor into voltages exactly proportional to the three components F_x , F_y and F_z of the acting force.

In order to construct multicomponent force plates for measuring three forces and three moments, special multichannel charge amplifiers are available.

Force Introduction

When only a single force link is used, then as far as possible, the center of pressure should be within the cover plate.

Eccentric force introduction produces a moment of force on the sensor element and is permitted only up to specified values. If such a moment prevails, then the areas of force must be reduced accordingly.

A rigidly constructed force plate with four force links largely prevents moment stresses on the sensor element.




Fig. 4: Mounting example: Drop test measurement

Systems for Multicomponent Measurements

Information concerning cable concept see data sheet for Multicomponent Measurements of Types 9017A, 9047B, 9067B and 9077B.

Optional Accessories Connecting cable Connecting cable 	Type 1693A 1694A
Ordering Key	Type
• 3-Component Force Link	9317B

Page 3/3

This information corresponds to the current state of knowledge. Kistler reserves the right to make technical changes. Liability for consequential damage resulting from the use of Kistler products is excluded.

©2004, Kistler Instrumente AG, PO Box, Eulachstr. 22, CH-8408 Winterthur Tel +41 52 224 11 11, Fax 224 14 14, info@kistler.com, www.kistler.com